

Scale type: PS 4500/X

Serial number:	393868	Minimal load:	0,5 g
Test date:	6/24/2013 12:23:27PM	Maximal capacity	4500 g
Scale class:	II	Verification division	0,1 g
Unit:	g	Read unit:	0,01 g
Working temperature: +15 °C +30 °C			

INDICATION ERROR


LOAD	0.5	500	2000	3000	4500	0	0	0	0	0
DEVIATION	-0.01	0	-0.01	0.01	-0.01					
D. ERROR	0.05	0.05	0.1	0.15	0.15	0	0	0	0	0

DISPERSION OF INDICATION

LOAD	2000	2000	2000	2000	2000
DEVIATION	-0.03	-0.02	-0.02	-0.02	-0.02
D. ERROR	0.1	0.1	0.1	0.1	0.1
LOAD	4500	4500	4500	4500	4500
DEVIATION	-0.02	-0.02	-0.02	-0.02	-0.02
D. ERROR	0.15	0.15	0.15	0.15	0.15

ECCENTRIC LOAD

	1	2	3	4
LOAD	1500	1500	1500	1500
DEVIATION	0.01	0.01	0.01	0.01
D. ERROR	0.1	0.1	0.1	0.1



Pan view

Temperature: °C
Pressure: %
Humidity: hPa

Scale has passed all the verification tests.

Data of verifying unit:

Radwag USA, LLC
USA, North Miami Beach, FL, 33179

Testing person:

Robert Sieminski

Print date: Monday, 24 June, 2013